

Location (tank or field)	Date	SeaFET ID	Average reads frame <sup>-1</sup>	Frames burst <sup>-1</sup>	Sampling freq. (min)	Calibration method
<i>APSH – Tank</i>	Oct. 5–8, 2016	395, 396, 397	1	10	5	factory
<i>OARC – Tank</i>	Oct. 26–Nov. 3, 2016	395, 396, 397	3	–	continuous	factory
<i>OARC – Tank</i>	Jan. 26–Feb. 1, 2017	395, 396, 397	1	10	180	factory
<i>APSH Field*</i>	Mar. 5–Jun. 6, 2017	397	10	30	180	factory, SP and MP in situ
<i>Kachemak Bay Field</i>	Mar. 18–Jun. 4, 2017	395, 396	10	30	180	factory, SP in situ
<i>Sentry Shoal Field</i>	Jul. 7–Aug. 24, Aug. 28–Nov. 29	268	10	30	30	factory, SP in situ